



URGENT: Field Safety Notice

Philips Respironics Ventilator Devices

Sound Abatement Foam
Susceptibility to Degradation and Volatile Organic Compound Emission

Cover Letter for Distributors and Institutions

Dear Device Customer,

Philips Respironics is issuing voluntarily Field Safety Corrective Action for Continuous and Non-Continuous Ventilators due to two (2) issues related to the polyester-based polyurethane (PE-PUR) sound abatement foam used in the design of these devices. Detailed information associated with this Field Safety Corrective Action and the impacted device populations can be found in the included Field Safety Notices.

You must take action, as a distributor or institution that is a direct customer of Philips, to isolate any impacted devices that are in your inventory. You must also facilitate the actions identified in the included Field Safety Notices for any impacted device users within your organization, among your past and present patient population, or among your customer base. You must take the following actions:

1. Read and understand the included Field Safety Notices. Distribute these notices to any member of your organization that must also be aware of the issues described.
2. Register for participation in this medical device Field Safety Corrective Action and receive additional instructions to participate in corrective actions by visiting the website:
www.philips.com/src-update

To respond to this Field Safety Notice, you will need your account number and passcode. This is included within the Response Instructions following the Field Safety Notices.

Philips regrets any inconveniences caused by this problem.

Sincerely,

Rodney Mell

Head of Quality and Regulatory

Philips Respironics - Sleep & Respiratory Care



URGENT: Field Safety Notice

Philips Respironics CPAP and Bi-Level PAP Devices

Sound Abatement Foam
Susceptibility to Degradation and Volatile Organic Compound Emission

Cover Letter for Distributors and Institutions

Dear Device Customer,

Philips Respironics is issuing voluntarily Field Safety Corrective Action for Continuous and Non-Continuous Ventilators due to two (2) issues related to the polyester-based polyurethane (PE-PUR) sound abatement foam used in the design of these devices. Detailed information associated with this Field Safety Corrective Action and the impacted device populations can be found in the included Field Safety Notices.

You must take action, as a distributor or institution that is a direct customer of Philips, to isolate any impacted devices that are in your inventory. You must also facilitate the actions identified in the included Field Safety Notices for any impacted device users within your organization, among your past and present patient population, or among your customer base. You must take the following actions:

1. Read and understand the included Field Safety Notices. Distribute these notices to any member of your organization that must also be aware of the issues described.
2. Register for participation in this medical device Field Safety Corrective Action and receive additional instructions to participate in corrective actions by visiting the website:
www.philips.com/src-update

To respond to this Field Safety Notice, you will need your account number and passcode. This is included within the Response Instructions following the Field Safety Notices.

Philips regrets any inconveniences caused by this problem.

Sincerely,

Rodney Mell

Head of Quality and Regulatory

Philips Respironics - Sleep & Respiratory Care